


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD070120\
Data File : VD065899.D
Acq On : 01 Jul 2020 09:39
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D062620S.M
Title : SW846 8260
Last Update : Fri Jun 26 19:16:07 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.2	5676	PASS
75	95	30	60	57.8	14780	PASS
95	95	100	100	100.0	25581	PASS
96	95	5	9	7.4	1903	PASS
173	174	0.00	2	1.5	310	PASS
174	95	50	100	83.6	21373	PASS
175	174	5	9	7.4	1579	PASS
176	174	95	101	95.5	20405	PASS
177	176	5	9	6.8	1379	PASS