

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070122\
 Data File : VD073763.D
 Acq On : 01 Jul 2022 14:36
 Operator : VA/SY
 Sample : N3502-07RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 D-44-UCRE

Quant Time: Jul 05 05:07:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D063022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 05 04:59:39 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	48425	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	99538	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	91666	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	40119	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	36446	77.150	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	154.300%
35) Dibromofluoromethane	7.909	113	32806	60.670	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	121.340%
50) Toluene-d8	10.332	98	123999	62.515	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	125.020%
62) 4-Bromofluorobenzene	12.620	95	43177	55.190	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.380%
Target Compounds						Qvalue
16) Acetone	4.156	43	5975	60.792	ug/l	99
25) 2-Butanone	7.215	43	1410	8.736	ug/l #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070122\
Data File : VD073763.D
Acq On : 01 Jul 2022 14:36
Operator : VA/SY
Sample : N3502-07RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
D-44-UCRE

Quant Time: Jul 05 05:07:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D063022S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 05 04:59:39 2022
Response via : Initial Calibration

