

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070122\
 Data File : VD073765.D
 Acq On : 01 Jul 2022 15:37
 Operator : VA/SY
 Sample : N3562-01
 Misc : 5.61G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-7A

Quant Time: Jul 05 05:08:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D063022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 05 04:59:39 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	171377	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	348715	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	310472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	134764	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	117448	70.251	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.500%
35) Dibromofluoromethane	7.908	113	109706	57.912	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.820%
50) Toluene-d8	10.332	98	431459	62.090	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	124.180%
62) 4-Bromofluorobenzene	12.620	95	151449	55.258	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.520%
Target Compounds						
						Qvalue
16) Acetone	4.156	43	24886	71.545	ug/l	91
17) Carbon Disulfide	4.403	76	7647	3.468	ug/l #	91
25) 2-Butanone	7.208	43	8445	14.784	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070122\
Data File : VD073765.D
Acq On : 01 Jul 2022 15:37
Operator : VA/SY
Sample : N3562-01
Misc : 5.61G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-7A

Quant Time: Jul 05 05:08:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D063022S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 05 04:59:39 2022
Response via : Initial Calibration

