

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070522\
 Data File : VD073778.D
 Acq On : 05 Jul 2022 17:06
 Operator : VA/SY
 Sample : N3562-07
 Misc : 3.64G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-10A

Quant Time: Jul 06 02:51:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D063022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 05 10:05:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	124360	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.855	114	250694	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	216009	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	78134	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	71004	58.528	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.060%
35) Dibromofluoromethane	7.903	113	71356	52.395	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.800%
50) Toluene-d8	10.332	98	206807	41.397	ug/l	0.00
Spiked Amount	50.000	Range	78 - 12	Recovery	=	82.800%#
62) 4-Bromofluorobenzene	12.620	95	80282	40.745	ug/l	0.00
Spiked Amount	50.000	Range	50 - 14	Recovery	=	81.480%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070522\
Data File : VD073778.D
Acq On : 05 Jul 2022 17:06
Operator : VA/SY
Sample : N3562-07
Misc : 3.64G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-10A

Quant Time: Jul 06 02:51:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D063022S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 05 10:05:34 2022
Response via : Initial Calibration

