

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070522\
 Data File : VD073779.D
 Acq On : 05 Jul 2022 17:35
 Operator : VA/SY
 Sample : N3562-09RE
 Misc : 6.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
ClientSampleId :
 SB-11ARE

Quant Time: Jul 06 02:51:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D063022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 05 10:05:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	106958	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.855	114	217350	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	152265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	36997	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	67647	64.833	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.660%
35) Dibromofluoromethane	7.908	113	66018	55.913	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.820%
50) Toluene-d8	10.332	98	165438	38.197	ug/l	0.00
Spiked Amount	50.000	Range	78 - 12	Recovery	=	76.400%#
62) 4-Bromofluorobenzene	12.620	95	51487	30.139	ug/l	0.00
Spiked Amount	50.000	Range	50 - 14	Recovery	=	60.280%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070522\
 Data File : VD073779.D
 Acq On : 05 Jul 2022 17:35
 Operator : VA/SY
 Sample : N3562-09RE
 Misc : 6.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-11ARE

Quant Time: Jul 06 02:51:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D063022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 05 10:05:34 2022
 Response via : Initial Calibration

