

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070522\
 Data File : VD073783.D
 Acq On : 05 Jul 2022 19:35
 Operator : VA/SY
 Sample : N3562-12
 Misc : 7.72G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-12B

Quant Time: Jul 06 02:52:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D063022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 05 10:05:34 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	107461	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	216452	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	189677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	66028	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	63236	60.321	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 120.640%	
35) Dibromofluoromethane	7.909	113	61682	52.457	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.920%	
50) Toluene-d8	10.332	98	175657	40.725	ug/l	0.00
Spiked Amount	50.000	Range	78 - 12	Recovery	= 81.440%#	
62) 4-Bromofluorobenzene	12.620	95	69410	40.800	ug/l	0.00
Spiked Amount	50.000	Range	50 - 14	Recovery	= 81.600%#	
Target Compounds						
					Qvalue	
16) Acetone	4.150	43	17288	79.264	ug/l	96
17) Carbon Disulfide	4.403	76	6538	4.729	ug/l	96
25) 2-Butanone	7.220	43	4604	12.854	ug/l	97
52) Toluene	10.397	92	5906	1.836	ug/l	94
68) m/p-Xylenes	11.838	106	3125	1.278	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070522\
Data File : VD073783.D
Acq On : 05 Jul 2022 19:35
Operator : VA/SY
Sample : N3562-12
Misc : 7.72G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-12B

Quant Time: Jul 06 02:52:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D063022S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 05 10:05:34 2022
Response via : Initial Calibration

