

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072022\
 Data File : VD073877.D
 Acq On : 20 Jul 2022 13:28
 Operator : VA/SY
 Sample : N3693-01
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-230

Quant Time: Jul 20 16:29:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 20 15:29:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	128368	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	256205	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	249741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	113876	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	83654	54.574	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.140%
35) Dibromofluoromethane	7.903	113	89845	49.707	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.420%
50) Toluene-d8	10.326	98	322646	49.172	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.340%
62) 4-Bromofluorobenzene	12.620	95	110377	48.926	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072022\
Data File : VD073877.D
Acq On : 20 Jul 2022 13:28
Operator : VA/SY
Sample : N3693-01
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-230

Quant Time: Jul 20 16:29:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 20 15:29:08 2022
Response via : Initial Calibration

