

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072022\
 Data File : VD073882.D
 Acq On : 20 Jul 2022 16:32
 Operator : VA/SY
 Sample : N3730-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-01-6.5-7.0

Quant Time: Jul 21 01:10:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 20 15:29:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	119097	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	244058	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	232929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	104141	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	77592	54.559	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.120%
35) Dibromofluoromethane	7.903	113	82824	48.104	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.200%
50) Toluene-d8	10.332	98	301878	48.297	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.600%
62) 4-Bromofluorobenzene	12.620	95	100036	46.549	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.100%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.991	85	5369	2.982	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072022\
Data File : VD073882.D
Acq On : 20 Jul 2022 16:32
Operator : VA/SY
Sample : N3730-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-01-6.5-7.0

Quant Time: Jul 21 01:10:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 20 15:29:08 2022
Response via : Initial Calibration

