

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072122\
 Data File : VD073895.D
 Acq On : 21 Jul 2022 14:27
 Operator : VA/SY
 Sample : N3810-03
 Misc : 5.25G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COMP-3

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 07/22/2022
 Supervised By :Mahesh Dadoda 07/22/2022

Quant Time: Jul 22 01:19:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 20 15:29:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	124410	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	246238	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	238550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	105651	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	77568	52.213	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 104.420%			
35) Dibromofluoromethane	7.903	113	86207	49.625	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 99.260%			
50) Toluene-d8	10.326	98	307071	48.693	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 97.380%			
62) 4-Bromofluorobenzene	12.620	95	105666	48.734	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 97.460%			
Target Compounds					Qvalue	
16) Acetone	4.156	43	29856	122.476	ug/l	93
17) Carbon Disulfide	4.409	76	4615	1.142	ug/l #	91
25) 2-Butanone	7.203	43	5300m	15.318	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072122\
Data File : VD073895.D
Acq On : 21 Jul 2022 14:27
Operator : VA/SY
Sample : N3810-03
Misc : 5.25G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
COMP-3

Quant Time: Jul 22 01:19:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 20 15:29:08 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 07/22/2022
Supervised By :Mahesh Dadoda 07/22/2022

