

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072122\
 Data File : VD073899.D
 Acq On : 21 Jul 2022 16:32
 Operator : VA/SY
 Sample : N3814-04
 Misc : 5.68G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ENV-DE-3(0-5)

Quant Time: Jul 22 01:19:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 20 15:29:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	127361	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	236325	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	224113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	96611	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	80051	52.636	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.280%
35) Dibromofluoromethane	7.909	113	82948	49.752	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.500%
50) Toluene-d8	10.332	98	294697	48.691	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.380%
62) 4-Bromofluorobenzene	12.620	95	97520	46.863	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072122\
Data File : VD073899.D
Acq On : 21 Jul 2022 16:32
Operator : VA/SY
Sample : N3814-04
Misc : 5.68G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ENV-DE-3(0-5)

Quant Time: Jul 22 01:19:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 20 15:29:08 2022
Response via : Initial Calibration

