

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072122\
 Data File : VD073901.D
 Acq On : 21 Jul 2022 17:34
 Operator : VA/SY
 Sample : N3814-06
 Misc : 6.37G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ENV-DE-3(10-15)

Quant Time: Jul 22 01:20:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 20 15:29:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	125612	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	232727	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	230671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	105819	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	81551	54.369	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.740%
35) Dibromofluoromethane	7.909	113	81035	49.356	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.720%
50) Toluene-d8	10.332	98	290348	48.714	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.420%
62) 4-Bromofluorobenzene	12.620	95	106403	51.923	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072122\
Data File : VD073901.D
Acq On : 21 Jul 2022 17:34
Operator : VA/SY
Sample : N3814-06
Misc : 6.37G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ENV-DE-3(10-15)

Quant Time: Jul 22 01:20:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 20 15:29:08 2022
Response via : Initial Calibration

