

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072122\
 Data File : VD073903.D
 Acq On : 21 Jul 2022 18:37
 Operator : VA/SY
 Sample : N3814-08
 Misc : 6.17G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ENV-DE-2(5-10)

Quant Time: Jul 22 01:20:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 20 15:29:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	124059	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	232798	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	219158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	97522	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	76810	51.849	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.700%
35) Dibromofluoromethane	7.908	113	80784	49.188	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.380%
50) Toluene-d8	10.332	98	288292	48.354	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.700%
62) 4-Bromofluorobenzene	12.620	95	97328	47.480	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072122\
Data File : VD073903.D
Acq On : 21 Jul 2022 18:37
Operator : VA/SY
Sample : N3814-08
Misc : 6.17G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ENV-DE-2(5-10)

Quant Time: Jul 22 01:20:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 20 15:29:08 2022
Response via : Initial Calibration

