

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072122\
 Data File : VD073904.D
 Acq On : 21 Jul 2022 19:08
 Operator : VA/SY
 Sample : N3814-09
 Misc : 6.32G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ENV-DE-2(10-15)

Quant Time: Jul 22 01:20:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 20 15:29:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	126067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	231345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	223258	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	102326	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	81570	54.185	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.380%
35) Dibromofluoromethane	7.908	113	81589	49.990	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.980%
50) Toluene-d8	10.332	98	289427	48.850	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.700%
62) 4-Bromofluorobenzene	12.620	95	102064	50.103	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072122\
Data File : VD073904.D
Acq On : 21 Jul 2022 19:08
Operator : VA/SY
Sample : N3814-09
Misc : 6.32G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ENV-DE-2(10-15)

Quant Time: Jul 22 01:20:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 20 15:29:08 2022
Response via : Initial Calibration

