

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072122\
 Data File : VD073890.D
 Acq On : 21 Jul 2022 11:55
 Operator : VA/SY
 Sample : VD0721SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0721SBL01

Quant Time: Jul 22 01:18:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 20 15:29:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	135801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	261604	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	253214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	115277	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	83876	51.724	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.440%
35) Dibromofluoromethane	7.903	113	91214	49.423	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.840%
50) Toluene-d8	10.332	98	326986	48.805	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.620%
62) 4-Bromofluorobenzene	12.620	95	111673	48.479	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.960%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072122\
Data File : VD073890.D
Acq On : 21 Jul 2022 11:55
Operator : VA/SY
Sample : VD0721SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0721SBL01

Quant Time: Jul 22 01:18:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 20 15:29:08 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.973 min Scan# 1028
Delta R.T. 0.006 min
Lab File: VD073890.D
Acq: 21 Jul 2022 11:55




Instrument :
MSVOA_D
ClientSampleId :
VD0721SBL01


Tgt Ion:168 Resp: 135801
Ion Ratio Lower Upper
168 100
99 58.4 49.8 74.6



#33
1,2-Dichloroethane-d4
Concen: 51.724 ug/l
RT: 8.320 min Scan# 1088
Delta R.T. -0.006 min
Lab File: VD073890.D
Acq: 21 Jul 2022 11:55

Tgt Ion: 65 Resp: 83876
Ion Ratio Lower Upper
65 100
67 51.0 0.0 101.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.855 min Scan# 1179

Delta R.T. -0.000 min

Lab File: VD073890.D

Acq: 21 Jul 2022 11:55

Instrument :

MSVOA_D

ClientSampleId :

VD0721SBL01





Tgt Ion:114 Resp: 261604

Ion Ratio Lower Upper

114 100

63 18.3 0.0 46.6

88 13.8 0.0 33.2



#35

Dibromofluoromethane

Concen: 49.423 ug/l

RT: 7.903 min Scan# 1017

Delta R.T. -0.006 min

Lab File: VD073890.D

Acq: 21 Jul 2022 11:55




Tgt Ion:113 Resp: 91214


Ion Ratio Lower Upper

113 100

111 101.7 83.8 125.6

192 17.9 14.6 21.8









#50
Toluene-d8
Concen: 48.805 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VD073890.D
Acq: 21 Jul 2022 11:55




Instrument :
MSVOA_D
ClientSampleId :
VD0721SBL01


Tgt Ion: 98 Resp: 326986
Ion Ratio Lower Upper
98 100
100 65.6 52.4 78.6



#62
4-Bromofluorobenzene
Concen: 48.479 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. -0.000 min
Lab File: VD073890.D
Acq: 21 Jul 2022 11:55

Tgt Ion: 95 Resp: 111673
Ion Ratio Lower Upper
95 100
174 80.5 0.0 152.8
176 77.4 0.0 145.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.632 min Scan# 1651
Delta R.T. -0.000 min
Lab File: VD073890.D
Acq: 21 Jul 2022 11:55

Instrument :
MSVOA_D
ClientSampleId :
VD0721SBL01

Tgt Ion:117 Resp: 253214
Ion Ratio Lower Upper
117 100
82 56.3 45.3 67.9
119 31.8 26.6 40.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. -0.000 min
Lab File: VD073890.D
Acq: 21 Jul 2022 11:55

Tgt Ion:152 Resp: 115277
Ion Ratio Lower Upper
152 100
115 61.0 31.3 93.8
150 160.4 0.0 348.6

