

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD072320\
 Data File : VD066025.D
 Acq On : 23 Jul 2020 15:35
 Operator : VA/SY
 Sample : L3381-07
 Misc : 6.56G/5.00ml/MSVOA D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 364-GRID-A7-A8

Quant Time: Jul 24 04:13:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072120S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 12:34:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	303530	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	485849	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	504867	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	251684	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	125115	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
35) Dibromofluoromethane	7.92	113	146741	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
50) Toluene-d8	10.34	98	586099	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
62) 4-Bromofluorobenzene	12.64	95	235758	61.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.88%	
Target Compounds						
16) Acetone	4.16	43	30871	61.752	ug/l	96
17) Carbon Disulfide	4.41	76	40471	4.972	ug/l	100
20) Methylene Chloride	4.97	84	24832	6.398	ug/l	93
25) 2-Butanone	7.21	43	11423	17.220	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD072320\
Data File : VD066025.D
Acq On : 23 Jul 2020 15:35
Operator : VA/SY
Sample : L3381-07
Misc : 6.56G/5.00ml/MSVOA D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
364-GRID-A7-A8

Quant Time: Jul 24 04:13:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D072120S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 12:34:13 2020
Response via : Initial Calibration

