

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD072320\
 Data File : VD066036.D
 Acq On : 23 Jul 2020 20:36
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 24 04:48:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072120S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 12:34:13 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	73	0.00
2 T	Dichlorodifluoromethane	0.379	0.330	12.9	68	0.00
3 P	Chloromethane	0.310	0.298	3.9	73	0.00
4 C	Vinyl Chloride	0.295	0.273	7.5#	68	0.00
5 T	Bromomethane	0.234	0.216	7.7	68	0.00
6 T	Chloroethane	0.176	0.181	-2.8	73	0.00
7 T	Trichlorofluoromethane	0.717	0.690	3.8	70	0.00
8 T	Diethyl Ether	0.198	0.217	-9.6	80	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.433	0.415	4.2	70	0.00
10 T	Methyl Iodide	0.324	0.508	-56.8#	99	0.00
11 T	Tert butyl alcohol	0.033	0.026	21.2#	72	-0.01
12 CM	1,1-Dichloroethene	0.413	0.403	2.4#	70	0.00
13 T	Acrolein	0.034	0.033	2.9	80	0.00
14 T	Allyl chloride	0.646	0.664	-2.8	72	0.00
15 T	Acrylonitrile	0.085	0.093	-9.4	79	0.00
16 T	Acetone	0.082	0.071	13.4	60	0.00
17 T	Carbon Disulfide	1.341	1.264	5.7	68	0.00
18 T	Methyl Acetate	0.192	0.188	2.1	75	0.00
19 T	Methyl tert-butyl Ether	0.849	0.990	-16.6	83	0.00
20 T	Methylene Chloride	0.521	0.532	-2.1	84	0.00
21 T	trans-1,2-Dichloroethene	0.467	0.489	-4.7	76	0.00
22 T	Diisopropyl ether	1.228	1.381	-12.5	79	0.00
23 T	Vinyl Acetate	0.692	0.791	-14.3	79	0.00
24 P	1,1-Dichloroethane	0.744	0.829	-11.4	79	0.00
25 T	2-Butanone	0.109	0.111	-1.8	72	0.00
26 T	2,2-Dichloropropane	0.669	0.670	-0.1	71	0.00
27 T	cis-1,2-Dichloroethene	0.495	0.541	-9.3	78	0.00
28 T	Bromochloromethane	0.316	0.361	-14.2	90	0.00
29 T	Tetrahydrofuran	0.067	0.072	-7.5	76	0.00
30 C	Chloroform	0.778	0.892	-14.7#	83	0.00
31 T	Cyclohexane	0.732	0.655	10.5	66	0.00
32 T	1,1,1-Trichloroethane	0.720	0.786	-9.2	79	0.00
33 S	1,2-Dichloroethane-d4	0.407	0.500	-22.9#	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	78	0.00
35 S	Dibromofluoromethane	0.303	0.342	-12.9	101	0.00
36 T	1,1-Dichloropropene	0.427	0.416	2.6	74	0.00
37 T	Ethyl Acetate	0.163	0.162	0.6	77	0.00
38 T	Carbon Tetrachloride	0.463	0.463	0.0	76	0.00
39 T	Methylcyclohexane	0.517	0.492	4.8	70	0.00
40 TM	Benzene	1.194	1.247	-4.4	79	0.00
41 T	Methacrylonitrile	0.093	0.119	-28.0#	95	-0.01
42 TM	1,2-Dichloroethane	0.327	0.362	-10.7	85	0.00
43 T	Isopropyl Acetate	0.318	0.330	-3.8	78	0.00
44 TM	Trichloroethene	0.361	0.377	-4.4	80	0.00
45 C	1,2-Dichloropropane	0.289	0.309	-6.9#	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD072320\
 Data File : VD066036.D
 Acq On : 23 Jul 2020 20:36
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 24 04:48:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072120S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 21 12:34:13 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.156	0.170	-9.0	83	0.00
47 T	Bromodichloromethane	0.412	0.455	-10.4	85	0.00
48 T	Methyl methacrylate	0.161	0.161	0.0	75	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	74	0.00
50 S	Toluene-d8	1.130	1.272	-12.6	96	0.00
51 T	4-Methyl-2-Pentanone	0.160	0.170	-6.3	79	0.00
52 CM	Toluene	0.779	0.840	-7.8#	80	0.00
53 T	t-1,3-Dichloropropene	0.383	0.426	-11.2	84	0.00
54 T	cis-1,3-Dichloropropene	0.465	0.514	-10.5	83	0.00
55 T	1,1,2-Trichloroethane	0.221	0.233	-5.4	81	0.00
56 T	Ethyl methacrylate	0.257	0.293	-14.0	82	0.00
57 T	1,3-Dichloropropane	0.365	0.409	-12.1	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.123	0.121	1.6	85	0.00
59 T	2-Hexanone	0.111	0.114	-2.7	75	0.00
60 T	Dibromochloromethane	0.298	0.334	-12.1	87	0.00
61 T	1,2-Dibromoethane	0.213	0.240	-12.7	86	0.00
62 S	4-Bromofluorobenzene	0.392	0.447	-14.0	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.339	0.351	-3.5	83	0.00
65 PM	Chlorobenzene	0.915	0.984	-7.5	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.338	0.375	-10.9	87	0.00
67 C	Ethyl Benzene	1.575	1.674	-6.3#	83	0.00
68 T	m/p-Xylenes	0.606	0.647	-6.8	83	0.00
69 T	o-Xylene	0.551	0.616	-11.8	86	0.00
70 T	Styrene	0.946	1.066	-12.7	87	0.00
71 P	Bromoform	0.198	0.213	-7.6	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	3.025	3.207	-6.0	84	0.00
74 T	N-amyl acetate	0.627	0.651	-3.8	82	0.00
75 P	1,1,2,2-Tetrachloroethane	0.500	0.516	-3.2	86	0.00
76 T	1,2,3-Trichloropropane	0.376	0.374	0.5	86	0.00
77 T	Bromobenzene	0.773	0.819	-6.0	88	0.00
78 T	n-propylbenzene	3.511	3.671	-4.6	82	0.00
79 T	2-Chlorotoluene	2.006	2.101	-4.7	84	0.00
80 T	1,3,5-Trimethylbenzene	2.486	2.717	-9.3	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.164	0.167	-1.8	84	0.00
82 T	4-Chlorotoluene	2.117	2.243	-6.0	85	0.00
83 T	tert-Butylbenzene	2.175	2.350	-8.0	84	0.00
84 T	1,2,4-Trimethylbenzene	2.452	2.719	-10.9	88	0.00
85 T	sec-Butylbenzene	2.996	3.113	-3.9	82	0.00
86 T	p-Isopropyltoluene	2.722	2.945	-8.2	84	0.00
87 T	1,3-Dichlorobenzene	1.436	1.545	-7.6	89	0.00
88 T	1,4-Dichlorobenzene	1.451	1.532	-5.6	88	0.00
89 T	n-Butylbenzene	2.531	2.615	-3.3	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD072320\
Data File : VD066036.D
Acq On : 23 Jul 2020 20:36
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jul 24 04:48:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072120S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 21 12:34:13 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.563	0.600	-6.6	88	0.00
91 T	1,2-Dichlorobenzene	1.244	1.353	-8.8	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.082	0.083	-1.2	86	0.00
93 T	1,2,4-Trichlorobenzene	0.893	0.978	-9.5	90	0.00
94 T	Hexachlorobutadiene	0.552	0.570	-3.3	84	0.00
95 T	Naphthalene	1.425	1.615	-13.3	90	0.00
96 T	1,2,3-Trichlorobenzene	0.776	0.868	-11.9	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6