


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD072320\
 Data File : VD066014.D
 Acq On : 23 Jul 2020 09:33
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072120S.M
 Title : SW846 8260
 Last Update : Tue Jul 21 12:34:13 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	28746	PASS
75	95	30	60	48.9	77328	PASS
95	95	100	100	100.0	158080	PASS
96	95	5	9	6.4	10112	PASS
173	174	0.00	2	0.4	635	PASS
174	95	50	100	91.2	144234	PASS
175	174	5	9	8.1	11700	PASS
176	174	95	101	95.7	138018	PASS
177	176	5	9	7.5	10343	PASS