

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD080720\
 Data File : VD066152.D
 Acq On : 07 Aug 2020 12:39
 Operator : VA/SY
 Sample : L3555-02
 Misc : 5.02G/5.00ml/MSVOA D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHRT-22546

Quant Time: Aug 08 01:01:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:49:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	240177	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	438717	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	461558	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	203960	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	124150	57.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.88%	
35) Dibromofluoromethane	7.91	113	139628	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
50) Toluene-d8	10.34	98	522694	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
62) 4-Bromofluorobenzene	12.64	95	202009	54.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.28%	
Target Compounds						
8) Diethyl Ether	3.70	74	6680	6.231	ug/l	87
16) Acetone	4.15	43	55816	137.285	ug/l	91
91) 1,2-Dichlorobenzene	13.90	146	6013	1.032	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD080720\
Data File : VD066152.D
Acq On : 07 Aug 2020 12:39
Operator : VA/SY
Sample : L3555-02
Misc : 5.02G/5.00ml/MSVOA D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHRT-22546

Quant Time: Aug 08 01:01:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:49:25 2020
Response via : Initial Calibration

