


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD080720\
 Data File : VD066147.D
 Acq On : 07 Aug 2020 09:31
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Title : SW846 8260
 Last Update : Wed Aug 05 13:49:25 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	26210	PASS
75	95	30	60	48.9	68133	PASS
95	95	100	100	100.0	139226	PASS
96	95	5	9	6.7	9393	PASS
173	174	0.00	2	0.5	674	PASS
174	95	50	100	91.2	126941	PASS
175	174	5	9	7.6	9630	PASS
176	174	95	101	95.6	121397	PASS
177	176	5	9	7.4	8996	PASS