

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081120\
 Data File : VD066192.D
 Acq On : 11 Aug 2020 17:50
 Operator : VA/SY
 Sample : L3624-12
 Misc : 3.91G/5.00ml/MSVOA D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SS-TP-3

Quant Time: Aug 11 21:09:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:49:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.97	168	183834	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	315029	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	390370	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	222521	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	91903	56.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.08%	
35) Dibromofluoromethane	7.91	113	103568	51.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.80%	
50) Toluene-d8	10.34	98	397270	53.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.88%	
62) 4-Bromofluorobenzene	12.63	95	189250	71.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	142.58%	
Target Compounds						
16) Acetone	4.16	43	8461	27.189	ug/l	96
20) Methylene Chloride	4.96	84	12979	2.865	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081120\
Data File : VD066192.D
Acq On : 11 Aug 2020 17:50
Operator : VA/SY
Sample : L3624-12
Misc : 3.91G/5.00ml/MSVOA D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SS-TP-3

Quant Time: Aug 11 21:09:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:49:25 2020
Response via : Initial Calibration

