

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081120\
 Data File : VD066191.D
 Acq On : 11 Aug 2020 17:23
 Operator : VA/SY
 Sample : L3624-07
 Misc : 3.27G/5.00ml/MSVOA D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SS-TP-2

Quant Time: Aug 11 21:08:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:49:25 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	238174	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	408334	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	483721	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	279267	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	107465	50.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.14%	
35) Dibromofluoromethane	7.91	113	126866	49.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.08%	
50) Toluene-d8	10.34	98	513481	53.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.58%	
62) 4-Bromofluorobenzene	12.63	95	242625	70.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	141.02%	
Target Compounds						
16) Acetone	4.16	43	5132	12.729	ug/l	Qvalue 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081120\
Data File : VD066191.D
Acq On : 11 Aug 2020 17:23
Operator : VA/SY
Sample : L3624-07
Misc : 3.27G/5.00ml/MSVOA D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SS-TP-2

Quant Time: Aug 11 21:08:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:49:25 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1030

Delta R.T. 0.01 min

Lab File: VD066191.D

Acq: 11 Aug 2020 17:23

Instrument :

MSVOA_D

ClientSampleId :


SS-TP-2

Tot Ion:168 Resp: 238174

Ion Ratio Lower Upper

168 100

99 54.2 42.2 63.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC

7.98

120000

100000

80000



60000

40000

20000

0

Time-->



#16

Acetone

Concen: 12.729 ug/l

RT: 4.16 min Scan# 380

Delta R.T. 0.01 min

Lab File: VD066191.D


Acq: 11 Aug 2020 17:23

Tgt Ion: 43 Resp: 5132

Ion Ratio Lower Upper

43 100

58 35.3 24.7 37.1



Abundance Ion 43.00 (42.70 to 43.70): VDC

Ion 58.00 (57.70 to 58.70): VDC

4.16

2000


1500


1000

500

0

Time-->









#33
 1,2-Dichloroethane-d4
 Concen: 50.575 ug/l
 RT: 8.33 min Scan# 1089
 Delta R.T. -0.01 min
 Lab File: VD066191.D
 Acq: 11 Aug 2020 17:23




Instrument :
 MSVOA_D
 ClientSampleId :
 SS-TP-2


Tot Ion: 65 Resp: 107465
 Ion Ratio Lower Upper
 65 100
 67 52.4 0.0 104.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.86 min Scan# 1180
 Delta R.T. 0.00 min
 Lab File: VD066191.D
 Acq: 11 Aug 2020 17:23

Tgt Ion: 114 Resp: 408334
 Ion Ratio Lower Upper
 114 100
 63 18.5 0.0 38.6
 88 14.3 0.0 30.0









#35
 Dibromofluoromethane
 Concen: 49.044 ug/l
 RT: 7.91 min Scan# 1018
 Delta R.T. 0.00 min
 Lab File: VD066191.D
 Acq: 11 Aug 2020 17:23




Instrument :
 MSVOA_D
 ClientSampleId :
 SS-TP-2


Tot Ion:113 Resp: 126866
 Ion Ratio Lower Upper
 113 100
 111 103.6 80.7 121.1
 192 23.0 18.4 27.6



#50
 Toluene-d8
 Concen: 53.792 ug/l
 RT: 10.34 min Scan# 1431
 Delta R.T. 0.00 min
 Lab File: VD066191.D
 Acq: 11 Aug 2020 17:23

Tgt Ion: 98 Resp: 513481
 Ion Ratio Lower Upper
 98 100
 100 65.7 52.6 78.8





#62

4-Bromofluorobenzene

Concen: 70.513 ug/l

RT: 12.63 min Scan# 1821

Delta R.T. -0.01 min

Lab File: VD066191.D

Acq: 11 Aug 2020 17:23

Instrument :

MSVOA_D

ClientSampleId :

SS-TP-2


Tot Ion: 95 Resp: 242625

Ion Ratio Lower Upper

95 100

174 89.4 0.0 180.8

176 86.1 0.0 181.2





Abundance Ion 94.90 (94.60 to 95.60): VD066125.D (-1813) (-)

Ion 173.80 (173.50 to 174.50): VD066125.D (-1813) (-)

Ion 175.80 (175.50 to 176.50): VD066125.D (-1813) (-)

12.63

Time-->



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.64 min Scan# 1653

Delta R.T. 0.00 min

Lab File: VD066191.D

Acq: 11 Aug 2020 17:23


Tgt Ion: 117 Resp: 483721

Ion Ratio Lower Upper

117 100

82 55.0 43.0 64.4

119 33.3 26.1 39.1




Abundance Ion 116.90 (116.60 to 117.60): VD066125.D (-1644) (-)


Ion 82.00 (81.70 to 82.70): VD066125.D (-1644) (-)

Ion 118.90 (118.60 to 119.60): VD066125.D (-1644) (-)

11.64

Time-->





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.58 min Scan# 1982

Delta R.T. 0.00 min

Lab File: VD066191.D

Acq: 11 Aug 2020 17:23

Instrument :

MSVOA_D

ClientSampleId :

SS-TP-2


Tot Ion:152 Resp: 279267

Ion Ratio Lower Upper

152 100

115 57.0 28.0 83.9

150 157.7 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

