

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081319\
 Data File : VD063538.D
 Acq On : 13 Aug 2019 20:16
 Operator : JC/SY
 Sample : K4319-07
 Misc : 6.25g/5ml/MSVOA D/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TEST-PIT-2

Quant Time: Aug 14 09:21:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	987565	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	8.22	114	1517923	50.00	ug/l	-0.02
63) Chlorobenzene-d5	12.36	117	1245733	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	598280	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	445970	42.25	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	84.50%	
35) Dibromofluoromethane	6.92	113	508572	39.78	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	79.56%	
50) Toluene-d8	10.41	98	1006347	35.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.62%	
62) 4-Bromofluorobenzene	13.56	95	389715	30.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	61.18%	
Target Compounds						
16) Acetone	3.22	43	67159	31.297	ug/l #	86
20) Methylene Chloride	3.81	84	88664	9.555	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081319\
Data File : VD063538.D
Acq On : 13 Aug 2019 20:16
Operator : JC/SY
Sample : K4319-07
Misc : 6.25g/5ml/MSVOA D/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TEST-PIT-2

Quant Time: Aug 14 09:21:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

