

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081319\
 Data File : VD063566.D
 Acq On : 14 Aug 2019 10:44
 Operator : JC/SY
 Sample : K4085-03
 Misc : 6.58g/5ml/MSVOA D/SOIL
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-05-080919-B

Quant Time: Aug 14 11:44:29 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	877142	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	8.22	114	1217565	50.00	ug/l	-0.01
63) Chlorobenzene-d5	12.38	117	714786	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	345063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	489847	52.24	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	104.48%	
35) Dibromofluoromethane	6.93	113	524324	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
50) Toluene-d8	10.42	98	752445	33.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.76%	
62) 4-Bromofluorobenzene	13.56	95	337259	33.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.02%	
Target Compounds						
16) Acetone	3.23	43	141355	74.165	ug/l	95
18) Methyl Acetate	3.64	43	17574	4.868	ug/l	95
20) Methylene Chloride	3.82	84	70999	8.615	ug/l	95
69) o-Xylene	13.06	106	53423	6.782	ug/l	96
80) 1,3,5-Trimethylbenzene	13.94	105	905512	47.219	ug/l	97
86) p-Isopropyltoluene	14.49	119	246034	11.999	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081319\
Data File : VD063566.D
Acq On : 14 Aug 2019 10:44
Operator : JC/SY
Sample : K4085-03
Misc : 6.58g/5ml/MSVOA D/SOIL
ALS Vial : 58 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HR-05-080919-B

Quant Time: Aug 14 11:44:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

