

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081622\
 Data File : VD074108.D
 Acq On : 16 Aug 2022 18:21
 Operator : VA/SY
 Sample : N4194-13
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PCB-GCP2-BLANK

Quant Time: Aug 16 20:53:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081522S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 01:41:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	132707	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	231928	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	199649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	78762	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	45517	45.365	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.720%
35) Dibromofluoromethane	7.909	113	60515	40.832	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.660%
50) Toluene-d8	10.332	98	143194	31.822	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	63.640%
62) 4-Bromofluorobenzene	12.620	95	59228	31.874	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	63.740%
Target Compounds						
16) Acetone	4.156	43	7048	18.396	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081622\
Data File : VD074108.D
Acq On : 16 Aug 2022 18:21
Operator : VA/SY
Sample : N4194-13
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PCB-GCP2-BLANK

Quant Time: Aug 16 20:53:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081522S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 01:41:25 2022
Response via : Initial Calibration

