

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081622\
 Data File : VD074111.D
 Acq On : 16 Aug 2022 19:47
 Operator : VA/SY
 Sample : N4189-15RE
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB12RE

Quant Time: Aug 16 20:54:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081522S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 01:41:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	27171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	50001	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	56750	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	26843	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	21830	118.246	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 236.500%#	
35) Dibromofluoromethane	7.909	113	22659	70.917	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 141.840%	
50) Toluene-d8	10.326	98	36639	37.768	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 75.540%	
62) 4-Bromofluorobenzene	12.620	95	19975	49.863	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 99.720%	
Target Compounds						
16) Acetone	4.162	43	3500	61.261	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081622\
Data File : VD074111.D
Acq On : 16 Aug 2022 19:47
Operator : VA/SY
Sample : N4189-15RE
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB12RE

Quant Time: Aug 16 20:54:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081522S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 01:41:25 2022
Response via : Initial Calibration

