

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081720\
 Data File : VD066267.D
 Acq On : 17 Aug 2020 19:32
 Operator : VA/SY
 Sample : L3691-03RE
 Misc : 8.81G/5.00ml/MSVOA D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S604-M-SO-1-1.5-081320RE

Quant Time: Aug 18 08:28:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:49:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	293911	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	473462	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	617669	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	368083	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	117017	44.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.26%	
35) Dibromofluoromethane	7.91	113	139022	46.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.70%	
50) Toluene-d8	10.34	98	596060	53.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.70%	
62) 4-Bromofluorobenzene	12.63	95	316923	79.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	158.88%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	4.09	101	99630	36.108	ug/l	98
16) Acetone	4.15	43	9855	19.808	ug/l #	78
18) Methyl Acetate	4.71	43	2990	2.366	ug/l	91
20) Methylene Chloride	4.95	84	32122	7.254	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081720\
Data File : VD066267.D
Acq On : 17 Aug 2020 19:32
Operator : VA/SY
Sample : L3691-03RE
Misc : 8.81G/5.00ml/MSVOA D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S604-M-SO-1-1.5-081320RE

Quant Time: Aug 18 08:28:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:49:25 2020
Response via : Initial Calibration

