

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081720\
 Data File : VD066268.D
 Acq On : 17 Aug 2020 20:00
 Operator : VA/SY
 Sample : L3691-04RE
 Misc : 7.95G/5.00ml/MSVOA D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S605-M-SO-1-1.5-081320RE

Quant Time: Aug 18 08:29:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:49:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.97	168	288666	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	464299	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	604467	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	369453	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	119499	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
35) Dibromofluoromethane	7.91	113	140178	47.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.32%	
50) Toluene-d8	10.34	98	585101	53.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.82%	
62) 4-Bromofluorobenzene	12.63	95	313317	80.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	160.16%	
Target Compounds						
16) Acetone	4.16	43	15095	30.891	ug/l	89
20) Methylene Chloride	4.96	84	56092	16.900	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081720\
Data File : VD066268.D
Acq On : 17 Aug 2020 20:00
Operator : VA/SY
Sample : L3691-04RE
Misc : 7.95G/5.00ml/MSVOA D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S605-M-SO-1-1.5-081320RE

Quant Time: Aug 18 08:29:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:49:25 2020
Response via : Initial Calibration

