

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081920\
 Data File : VD066319.D
 Acq On : 19 Aug 2020 15:30
 Operator : VA/SY
 Sample : L3700-08RE
 Misc : 8.87G/5.00ml/MSVOA D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S408-K-SO-1-1.5-081420RE

Quant Time: Aug 19 16:19:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:49:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	101640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	148789	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	113741	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	33805	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	96529	106.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	212.90%	
35) Dibromofluoromethane	7.91	113	96607	102.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	204.98%	
50) Toluene-d8	10.34	98	285490	82.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	164.16%	
62) 4-Bromofluorobenzene	12.64	95	68378	54.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.08%	
Target Compounds						
16) Acetone	4.16	43	10335	60.068	ug/l #	81
20) Methylene Chloride	4.97	84	12709	9.041	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081920\
Data File : VD066319.D
Acq On : 19 Aug 2020 15:30
Operator : VA/SY
Sample : L3700-08RE
Misc : 8.87G/5.00ml/MSVOA D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S408-K-SO-1-1.5-081420RE

Quant Time: Aug 19 16:19:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:49:25 2020
Response via : Initial Calibration

