

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081920\
 Data File : VD066336.D
 Acq On : 20 Aug 2020 01:25
 Operator : VA/SY
 Sample : L3712-04RE
 Misc : 6.74G/5.00ml/MSVOA D/SOIL
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHRT-26572RE

Quant Time: Aug 20 08:05:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:49:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	174466	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	279707	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	237372	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	99505	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	95388	61.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.56%	
35) Dibromofluoromethane	7.91	113	98040	55.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.66%	
50) Toluene-d8	10.34	98	296435	45.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.66%	
62) 4-Bromofluorobenzene	12.63	95	94779	40.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.42%	
Target Compounds						
8) Diethyl Ether	3.71	74	2328	2.990	ug/l	96
20) Methylene Chloride	4.97	84	40692	21.316	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081920\
Data File : VD066336.D
Acq On : 20 Aug 2020 01:25
Operator : VA/SY
Sample : L3712-04RE
Misc : 6.74G/5.00ml/MSVOA D/SOIL
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHRT-26572RE

Quant Time: Aug 20 08:05:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:49:25 2020
Response via : Initial Calibration

