

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081920\
 Data File : VD066326.D
 Acq On : 19 Aug 2020 18:40
 Operator : VA/SY
 Sample : L3712-08
 Misc : 9.42G/5.00ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VB-27207

Quant Time: Aug 20 07:47:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:49:25 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.97	168	155431	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	255451	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	204513	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	71654	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	86306	62.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.48%	
35) Dibromofluoromethane	7.91	113	89422	55.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.52%	
50) Toluene-d8	10.34	98	291643	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
62) 4-Bromofluorobenzene	12.63	95	81008	37.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.26%	
Target Compounds						
20) Methylene Chloride	4.97	84	19602	9.163	ug/l #	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081920\
Data File : VD066326.D
Acq On : 19 Aug 2020 18:40
Operator : VA/SY
Sample : L3712-08
Misc : 9.42G/5.00ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VB-27207

Quant Time: Aug 20 07:47:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:49:25 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.97 min Scan# 1029

Delta R.T. 0.00 min

Lab File: VD066326.D

Acq: 19 Aug 2020 18:40

Instrument :

MSVOA_D

ClientSampleId :


VB-27207

Tot Ion:168 Resp: 155431

Ion Ratio Lower Upper

168 100

99 49.5 42.2 63.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC

7.97

80000

60000



40000

20000

0

Time-->

7.90 8.00 8.10



#20

Methylene Chloride

Concen: 9.163 ug/l

RT: 4.97 min Scan# 518

Delta R.T. 0.01 min

Lab File: VD066326.D

Acq: 19 Aug 2020 18:40

Tgt Ion: 84 Resp: 19602


Ion Ratio Lower Upper

84 100

49 112.0 93.5 140.3

51 49.6 28.2 42.4#

86 54.4 49.5 74.3



Abundance Ion 83.90 (83.60 to 84.60): VDC

Ion 48.90 (48.60 to 49.60): VDC

Ion 50.90 (50.60 to 51.60): VDC

Ion 85.90 (85.60 to 86.60): VDC

10000

8000

6000


4000


2000

0

Time-->

4.90 5.00









#33
 1,2-Dichloroethane-d4
 Concen: 62.239 ug/l
 RT: 8.33 min Scan# 1089
 Delta R.T. -0.01 min
 Lab File: VD066326.D
 Acq: 19 Aug 2020 18:40




Instrument :
 MSVOA_D
 ClientSampleId :
 VB-27207


Tot Ion: 65 Resp: 86306
 Ion Ratio Lower Upper
 65 100
 67 48.9 0.0 104.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.86 min Scan# 1179
 Delta R.T. -0.01 min
 Lab File: VD066326.D
 Acq: 19 Aug 2020 18:40

Tgt Ion: 114 Resp: 255451
 Ion Ratio Lower Upper
 114 100
 63 19.5 0.0 38.6
 88 15.2 0.0 30.0









#35
 Dibromofluoromethane
 Concen: 55.258 ug/l
 RT: 7.91 min Scan# 1018
 Delta R.T. 0.00 min
 Lab File: VD066326.D
 Acq: 19 Aug 2020 18:40




Instrument :
 MSVOA_D
 ClientSampleId :
 VB-27207


Tot Ion:113 Resp: 89422
 Ion Ratio Lower Upper
 113 100
 111 104.3 80.7 121.1
 192 23.8 18.4 27.6



#50
 Toluene-d8
 Concen: 48.837 ug/l
 RT: 10.34 min Scan# 1431
 Delta R.T. 0.00 min
 Lab File: VD066326.D
 Acq: 19 Aug 2020 18:40

Tgt Ion: 98 Resp: 291643
 Ion Ratio Lower Upper
 98 100
 100 65.6 52.6 78.8





#62

4-Bromofluorobenzene

Concen: 37.633 ug/l

RT: 12.63 min Scan# 1821

Delta R.T. -0.01 min

Lab File: VD066326.D

Acq: 19 Aug 2020 18:40

Instrument :

MSVOA_D

ClientSampleId :

VB-27207


Tot Ion: 95 Resp: 81008

Ion Ratio Lower Upper

95 100

174 101.5 0.0 180.8




176 90.7 0.0 181.2



Abundance Ion 94.90 (94.60 to 95.60): V

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.64 min Scan# 1653

Delta R.T. 0.00 min

Lab File: VD066326.D

Acq: 19 Aug 2020 18:40


Tgt Ion: 117 Resp: 204513

Ion Ratio Lower Upper

117 100

82 53.6 43.0 64.4



119 34.3 26.1 39.1




Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): V

Ion 118.90 (118.60 to 119.60): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.58 min Scan# 1982

Delta R.T. 0.00 min

Lab File: VD066326.D

Acq: 19 Aug 2020 18:40

Instrument :

MSVOA_D

ClientSampleId :

VB-27207


Tot Ion:152 Resp: 71654

Ion Ratio Lower Upper

152 100

115 59.7 28.0 83.9

150 149.8 0.0 346.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

