

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081922\
 Data File : VD074152.D
 Acq On : 19 Aug 2022 16:58
 Operator : VA/SY
 Sample : N4215-01
 Misc : 5.94G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-1

Quant Time: Aug 20 01:25:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 19 09:25:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	168674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	297113	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	280263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	134418	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	66571	53.494	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.980%
35) Dibromofluoromethane	7.902	113	83852	58.263	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.520%
50) Toluene-d8	10.326	98	191571	46.718	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.440%
62) 4-Bromofluorobenzene	12.620	95	108906	48.986	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081922\
Data File : VD074152.D
Acq On : 19 Aug 2022 16:58
Operator : VA/SY
Sample : N4215-01
Misc : 5.94G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-1

Quant Time: Aug 20 01:25:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081822S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 19 09:25:06 2022
Response via : Initial Calibration

