

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081922\
 Data File : VD074156.D
 Acq On : 19 Aug 2022 18:55
 Operator : VA/SY
 Sample : N4227-01
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-1

Quant Time: Aug 20 01:27:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 19 09:25:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	135202	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	243277	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	230133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	97883	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	65755	67.509	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.020%
35) Dibromofluoromethane	7.909	113	81260	69.788	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	139.580%
50) Toluene-d8	10.332	98	174099	52.552	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.100%
62) 4-Bromofluorobenzene	12.620	95	91275	50.141	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.280%
Target Compounds						
16) Acetone	4.162	43	4041	13.873	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081922\
Data File : VD074156.D
Acq On : 19 Aug 2022 18:55
Operator : VA/SY
Sample : N4227-01
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-1

Quant Time: Aug 20 01:27:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081822S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 19 09:25:06 2022
Response via : Initial Calibration

