

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081922\
 Data File : VD074157.D
 Acq On : 19 Aug 2022 19:24
 Operator : VA/SY
 Sample : N4227-11
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-2

Quant Time: Aug 20 01:28:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 19 09:25:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	4363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	8123	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	9276	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	4055	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	3885	129.286	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	258.580%#
35) Dibromofluoromethane	7.897	113	3318	86.351	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	172.700%#
50) Toluene-d8	10.326	98	5963	54.070	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.140%
62) 4-Bromofluorobenzene	12.614	95	3697	60.824	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	121.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081922\
Data File : VD074157.D
Acq On : 19 Aug 2022 19:24
Operator : VA/SY
Sample : N4227-11
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-2

Quant Time: Aug 20 01:28:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081822S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 19 09:25:06 2022
Response via : Initial Calibration

