

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081922\
 Data File : VD074163.D
 Acq On : 19 Aug 2022 22:18
 Operator : VA/SY
 Sample : N4240-01
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-06-081722

Quant Time: Aug 20 01:30:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 19 09:25:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	140261	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	250284	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.631	117	226135	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	102396	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	67795	67.051	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.100%
35) Dibromofluoromethane	7.902	113	80889	67.377	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	134.760%
50) Toluene-d8	10.332	98	162571	47.111	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.220%
62) 4-Bromofluorobenzene	12.620	95	95142	50.802	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081922\
Data File : VD074163.D
Acq On : 19 Aug 2022 22:18
Operator : VA/SY
Sample : N4240-01
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-06-081722

Quant Time: Aug 20 01:30:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081822S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 19 09:25:06 2022
Response via : Initial Calibration

