

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082622\
 Data File : VD074236.D
 Acq On : 26 Aug 2022 14:30
 Operator : VA/SY
 Sample : N4310-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-02-082322

Quant Time: Aug 26 15:59:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 23 13:12:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	66608	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	124205	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	120735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	50827	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	75947	111.583	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 223.160%#	
35) Dibromofluoromethane	7.902	113	59330	71.414	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 142.820%	
50) Toluene-d8	10.332	98	157902	59.055	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 118.100%	
62) 4-Bromofluorobenzene	12.620	95	56269	53.270	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 106.540%	
Target Compounds						
16) Acetone	4.156	43	6072	46.825	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082622\
Data File : VD074236.D
Acq On : 26 Aug 2022 14:30
Operator : VA/SY
Sample : N4310-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-02-082322

Quant Time: Aug 26 15:59:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 23 13:12:17 2022
Response via : Initial Calibration

