

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082622\
 Data File : VD074237.D
 Acq On : 26 Aug 2022 14:59
 Operator : VA/SY
 Sample : N4326-01RE
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-03-082322RE

Quant Time: Aug 26 15:59:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 23 13:12:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	17762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	33013	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	34183	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	15292	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	21433	118.088	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	236.180%#
35) Dibromofluoromethane	7.908	113	16626	75.292	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	150.580%#
50) Toluene-d8	10.326	98	38978	54.846	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.700%
62) 4-Bromofluorobenzene	12.614	95	14887	53.024	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082622\
Data File : VD074237.D
Acq On : 26 Aug 2022 14:59
Operator : VA/SY
Sample : N4326-01RE
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-03-082322RE

Quant Time: Aug 26 15:59:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 23 13:12:17 2022
Response via : Initial Calibration

