



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082622\
Data File : VD074232.D
Acq On : 26 Aug 2022 10:10
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
Title : SW846 8260
Last Update : Tue Aug 23 13:12:17 2022



AutoFind: Scans 1817, 1818, 1819; Background Corrected with Scan 1809

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.0	9917	PASS
75	95	30	60	48.2	26576	PASS
95	95	100	100	100.0	55088	PASS
96	95	5	9	6.5	3606	PASS
173	174	0.00	2	1.5	652	PASS
174	95	50	100	78.0	42968	PASS
175	174	5	9	8.4	3606	PASS
176	174	95	101	98.1	42163	PASS
177	176	5	9	7.9	3318	PASS