

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD082820\
 Data File : VD066458.D
 Acq On : 28 Aug 2020 14:27
 Operator : VA/SY
 Sample : L3827-07
 Misc : 6.61G/5.00ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 789UMR-TP11-0006-01

Quant Time: Aug 29 01:19:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D082120S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 13:36:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.97	168	182452	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	259237	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	219392	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	92434	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	102321	57.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.40%	
35) Dibromofluoromethane	7.91	113	77572	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
50) Toluene-d8	10.34	98	290018	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
62) 4-Bromofluorobenzene	12.63	95	89721	42.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.26%	
Target Compounds						
68) m/p-Xylenes	11.86	106	9998	3.414	ug/l	87
69) o-Xylene	12.18	106	4366	1.603	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD082820\
Data File : VD066458.D
Acq On : 28 Aug 2020 14:27
Operator : VA/SY
Sample : L3827-07
Misc : 6.61G/5.00ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
789UMR-TP11-0006-01

Quant Time: Aug 29 01:19:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D082120S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 13:36:12 2020
Response via : Initial Calibration

