

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD091420\
 Data File : VD066690.D
 Acq On : 15 Sep 2020 00:43
 Operator : VA/SY
 Sample : L3981-10
 Misc : 11.76G/5.00ml/MSVOA D/SOIL
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B20-0-2

Quant Time: Sep 15 03:20:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D090320S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 04 01:57:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	362707	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	682801	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	649583	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	272751	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	243135	55.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.80%	
35) Dibromofluoromethane	7.91	113	237977	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
50) Toluene-d8	10.34	98	851792	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
62) 4-Bromofluorobenzene	12.63	95	276037	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
Target Compounds						
44) Trichloroethene	9.11	130	5776	1.131	ug/l	90
64) Tetrachloroethene	10.87	164	367482	84.022	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD091420\
Data File : VD066690.D
Acq On : 15 Sep 2020 00:43
Operator : VA/SY
Sample : L3981-10
Misc : 11.76G/5.00ml/MSVOA D/SOIL
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B20-0-2

Quant Time: Sep 15 03:20:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D090320S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 04 01:57:04 2020
Response via : Initial Calibration

