

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD091820\
 Data File : VD066761.D
 Acq On : 18 Sep 2020 12:56
 Operator : VA/SY
 Sample : L3870-06
 Misc : 5.21G/5.00ml/MSVOA D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-02-091720

Quant Time: Sep 19 03:47:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 15 16:35:37 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	76712	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	128793	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	145006	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	62296	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	74071	90.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	181.00%	
35) Dibromofluoromethane	7.91	113	58246	73.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	147.52%	
50) Toluene-d8	10.34	98	158887	56.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.96%	
62) 4-Bromofluorobenzene	12.63	95	57217	58.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.84%	
Target Compounds						
16) Acetone	4.16	43	20321	121.922	ug/l #	79
20) Methylene Chloride	4.96	84	11375	6.255	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD091820\
Data File : VD066761.D
Acq On : 18 Sep 2020 12:56
Operator : VA/SY
Sample : L3870-06
Misc : 5.21G/5.00ml/MSVOA D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-02-091720

Quant Time: Sep 19 03:47:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 15 16:35:37 2020
Response via : Initial Calibration

