

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD091820\
 Data File : VD066773.D
 Acq On : 18 Sep 2020 18:26
 Operator : VA/SY
 Sample : L4017-07
 Misc : 4.49G/5.00ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SS-7

Quant Time: Sep 19 06:42:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 15 16:35:37 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	335465	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	610111	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	613059	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	270997	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	218113	60.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	121.88%	
35) Dibromofluoromethane	7.91	113	201983	54.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.00%	
50) Toluene-d8	10.34	98	760338	57.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.10%	
62) 4-Bromofluorobenzene	12.63	95	260127	56.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.10%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD091820\
Data File : VD066773.D
Acq On : 18 Sep 2020 18:26
Operator : VA/SY
Sample : L4017-07
Misc : 4.49G/5.00ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SS-7

Quant Time: Sep 19 06:42:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 15 16:35:37 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1030

Delta R.T. -0.00 min

Lab File: VD066773.D

Acq: 18 Sep 2020 18:26

Instrument :

MSVOA_D

ClientSampleId :


SS-7

Tot Ion:168 Resp: 335465

Ion Ratio Lower Upper




168 100

99 63.6 54.4 81.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC



#33

1,2-Dichloroethane-d4

Concen: 60.939 ug/l

RT: 8.33 min Scan# 1090

Delta R.T. -0.00 min

Lab File: VD066773.D


Acq: 18 Sep 2020 18:26

Tgt Ion: 65 Resp: 218113

Ion Ratio Lower Upper



65 100


67 52.0 0.0 102.2



Abundance Ion 64.90 (64.60 to 65.60): VDC

Ion 66.90 (66.60 to 67.60): VDC









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.86 min Scan# 1180
 Delta R.T. -0.00 min
 Lab File: VD066773.D
 Acq: 18 Sep 2020 18:26




Instrument :
 MSVOA_D
 ClientSampleId :
 SS-7


Tot Ion	Ratio	Lower	Upper
114	100		
63	20.7	0.0	37.6
88	17.1	0.0	32.2



#35
 Dibromofluoromethane
 Concen: 53.998 ug/l
 RT: 7.91 min Scan# 1019
 Delta R.T. 0.01 min
 Lab File: VD066773.D
 Acq: 18 Sep 2020 18:26

Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.4	83.2	124.8
192	18.6	14.8	22.2





#50

Toluene-d8

Concen: 57.054 ug/l

RT: 10.34 min Scan# 1431

Delta R.T. -0.00 min

Lab File: VD066773.D

Acq: 18 Sep 2020 18:26

Instrument :

MSVOA_D

ClientSampleId :





SS-7

Tot Ion: 98 Resp: 760338

Ion Ratio Lower Upper

98 100

100 65.8 52.7 79.1



#62

4-Bromofluorobenzene

Concen: 56.547 ug/l

RT: 12.63 min Scan# 1821

Delta R.T. -0.00 min

Lab File: VD066773.D

Acq: 18 Sep 2020 18:26




Tgt Ion: 95 Resp: 260127


Ion Ratio Lower Upper

95 100

174 78.0 0.0 156.6

176 73.8 0.0 150.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.64 min Scan# 1653
Delta R.T. -0.00 min
Lab File: VD066773.D
Acq: 18 Sep 2020 18:26

Instrument :
MSVOA_D
ClientSampleId :
SS-7

Tot Ion	Ratio	Lower	Upper
117	100		
82	56.6	45.4	68.0
119	31.6	25.8	38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.58 min Scan# 1982
Delta R.T. -0.00 min
Lab File: VD066773.D
Acq: 18 Sep 2020 18:26

Tgt Ion	Ratio	Lower	Upper
152	100		
115	60.9	30.5	91.5
150	157.2	0.0	344.6

