

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD092120\
 Data File : VD066793.D
 Acq On : 21 Sep 2020 20:00
 Operator : VA/SY
 Sample : L4059-01
 Misc : 6.07G/5.00ml/MSVOA D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PAV-B33-091720-0-1

Quant Time: Sep 22 01:39:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 15 16:35:37 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	388522	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	729367	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	705946	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	303241	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	224543	54.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.34%	
35) Dibromofluoromethane	7.91	113	239798	53.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.26%	
50) Toluene-d8	10.34	98	887260	55.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.38%	
62) 4-Bromofluorobenzene	12.63	95	297482	54.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.18%	
Target Compounds						
20) Methylene Chloride	4.97	84	84854	13.295	ug/l	84
52) Toluene	10.40	92	35591	2.738	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD092120\
Data File : VD066793.D
Acq On : 21 Sep 2020 20:00
Operator : VA/SY
Sample : L4059-01
Misc : 6.07G/5.00ml/MSVOA D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PAV-B33-091720-0-1

Quant Time: Sep 22 01:39:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 15 16:35:37 2020
Response via : Initial Calibration

