


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD092120\
 Data File : VD066779.D
 Acq On : 21 Sep 2020 11:50
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D091520S.M
 Title : SW846 8260
 Last Update : Tue Sep 15 16:35:37 2020



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	39290	PASS
75	95	30	60	51.1	105376	PASS
95	95	100	100	100.0	206058	PASS
96	95	5	9	7.2	14775	PASS
173	174	0.00	2	1.6	2551	PASS
174	95	50	100	77.6	159930	PASS
175	174	5	9	7.4	11865	PASS
176	174	95	101	95.4	152517	PASS
177	176	5	9	7.3	11071	PASS