

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092721\
 Data File : VD070472.D
 Acq On : 27 Sep 2021 12:40
 Operator : VA/SY
 Sample : M3920-02RE
 Misc : 7.74G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-839-R-SO-2-2.5-092321RE

Quant Time: Sep 28 04:12:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	7915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	15214	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	14867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	7058	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	8713	85.732	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 171.460%#	
35) Dibromofluoromethane	7.909	113	5486	50.658	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.320%	
50) Toluene-d8	10.332	98	14982	38.103	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 76.200%	
62) 4-Bromofluorobenzene	12.626	95	6149	44.747	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.500%	
Target Compounds						Qvalue
16) Acetone	4.168	43	5030	277.217	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092721\
Data File : VD070472.D
Acq On : 27 Sep 2021 12:40
Operator : VA/SY
Sample : M3920-02RE
Misc : 7.74G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-839-R-SO-2-2.5-092321RE

Quant Time: Sep 28 04:12:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

