

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092721\
 Data File : VD070483.D
 Acq On : 27 Sep 2021 17:47
 Operator : VA/SY
 Sample : VIBLK
 Misc : 6.48G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Sep 28 04:15:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	357434	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	720216	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	755007	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	336664	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	246599	53.731	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.460%	
35) Dibromofluoromethane	7.914	113	251457	49.050	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 98.100%	
50) Toluene-d8	10.338	98	948715	50.969	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 101.940%	
62) 4-Bromofluorobenzene	12.620	95	356308	54.773	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 109.540%	
Target Compounds						
39) Methylcyclohexane	9.349	83	20713	2.408	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092721\
Data File : VD070483.D
Acq On : 27 Sep 2021 17:47
Operator : VA/SY
Sample : VIBLK
Misc : 6.48G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Sep 28 04:15:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

