

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100220\
 Data File : VD067064.D
 Acq On : 03 Oct 2020 05:39
 Operator : VA/SY
 Sample : L4247-10
 Misc : 5.02G/5.00ml/MSVOA D/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-20(12-14)

Quant Time: Oct 03 07:20:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 15:19:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.99	168	564613	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	924741	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	850835	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	385685	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	262198	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
35) Dibromofluoromethane	7.91	113	292107	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
50) Toluene-d8	10.34	98	1086256	49.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.48%	
62) 4-Bromofluorobenzene	12.63	95	363111	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
Target Compounds						
16) Acetone	4.16	43	13618	16.404	ug/l #	75
17) Carbon Disulfide	4.41	76	33181	2.031	ug/l	97
95) Naphthalene	15.41	128	40801	3.516	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD100220\
Data File : VD067064.D
Acq On : 03 Oct 2020 05:39
Operator : VA/SY
Sample : L4247-10
Misc : 5.02G/5.00ml/MSVOA D/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-20(12-14)

Quant Time: Oct 03 07:20:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 15:19:17 2020
Response via : Initial Calibration

