

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100220\
 Data File : VD067065.D
 Acq On : 03 Oct 2020 06:06
 Operator : VA/SY
 Sample : L4221-01
 Misc : 7.97G/5.00ml/MSVOA D/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-02-100120

Quant Time: Oct 03 07:21:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 15:19:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	553104	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	922058	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	826048	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	366869	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	261205	49.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.06%	
35) Dibromofluoromethane	7.91	113	291762	48.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.66%	
50) Toluene-d8	10.34	98	1080670	49.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.26%	
62) 4-Bromofluorobenzene	12.64	95	344811	46.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.36%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD100220\
Data File : VD067065.D
Acq On : 03 Oct 2020 06:06
Operator : VA/SY
Sample : L4221-01
Misc : 7.97G/5.00ml/MSVOA D/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-02-100120

Quant Time: Oct 03 07:21:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 15:19:17 2020
Response via : Initial Calibration

