

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100220\
 Data File : VD067059.D
 Acq On : 03 Oct 2020 03:21
 Operator : VA/SY
 Sample : L4247-05
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-18(11-13)

Quant Time: Oct 03 07:11:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 15:19:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	580444	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	958601	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	883925	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	407489	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.33	65	265846	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
35) Dibromofluoromethane	7.91	113	292623	46.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.26%	
50) Toluene-d8	10.34	98	1126441	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
62) 4-Bromofluorobenzene	12.63	95	388811	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD100220\
Data File : VD067059.D
Acq On : 03 Oct 2020 03:21
Operator : VA/SY
Sample : L4247-05
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-18(11-13)

Quant Time: Oct 03 07:11:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 15:19:17 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1030

Delta R.T. 0.00 min

Lab File: VD067059.D

Acq: 03 Oct 2020 03:21

Instrument :

MSVOA_D

ClientSampleId :


B-18(11-13)

Tot Ion:168 Resp: 580444

Ion Ratio Lower Upper




168 100

99 56.8 44.1 66.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC



#33

1,2-Dichloroethane-d4

Concen: 47.551 ug/l

RT: 8.33 min Scan# 1089

Delta R.T. -0.01 min

Lab File: VD067059.D


Acq: 03 Oct 2020 03:21

Tgt Ion: 65 Resp: 265846

Ion Ratio Lower Upper



65 100


67 52.0 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VDC

Ion 66.90 (66.60 to 67.60): VDC





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.86 min Scan# 1180

Delta R.T. 0.00 min

Lab File: VD067059.D

Acq: 03 Oct 2020 03:21

Instrument :

MSVOA_D

ClientSampleId :

B-18(11-13)


Tot Ion:114 Resp: 958601

Ion Ratio Lower Upper

114 100

63 17.9 0.0 38.4




88 15.7 0.0 31.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): V

Ion 87.90 (87.60 to 88.60): V



#35

Dibromofluoromethane

Concen: 46.627 ug/l

RT: 7.91 min Scan# 1019

Delta R.T. 0.00 min

Lab File: VD067059.D

Acq: 03 Oct 2020 03:21


Tgt Ion:113 Resp: 292623

Ion Ratio Lower Upper

113 100

111 104.2 83.0 124.4



192 20.8 15.5 23.3




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.258 ug/l

RT: 10.34 min Scan# 1431

Delta R.T. 0.00 min

Lab File: VD067059.D

Acq: 03 Oct 2020 03:21

Instrument :

MSVOA_D

ClientSampleId :


B-18(11-13)

Tot Ion: 98 Resp: 1126441




Ion Ratio Lower Upper

98 100

100 66.2 52.8 79.2



Abundance Ion 98.00 (97.70 to 98.70): VD066988.D (-1421) (-)



#62

4-Bromofluorobenzene

Concen: 50.084 ug/l

RT: 12.63 min Scan# 1821

Delta R.T. 0.00 min

Lab File: VD067059.D

Acq: 03 Oct 2020 03:21


Tgt Ion: 95 Resp: 388811

Ion Ratio Lower Upper



95 100


174 82.2 0.0 167.6

176 77.4 0.0 162.4



Abundance Ion 94.90 (94.60 to 95.60): VD066988.D (-1813) (-)









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.64 min Scan# 1653
Delta R.T. 0.00 min
Lab File: VD067059.D
Acq: 03 Oct 2020 03:21

Instrument :
MSVOA_D
ClientSampleId :
B-18(11-13)

Tot Ion	Ratio	Lower	Upper
117	100		
82	53.8	42.6	63.8
119	32.6	26.3	39.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.58 min Scan# 1982
Delta R.T. 0.00 min
Lab File: VD067059.D
Acq: 03 Oct 2020 03:21

Tgt Ion	Ratio	Lower	Upper
152	100		
115	58.3	29.6	88.8
150	155.4	0.0	344.8

