

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100520\
 Data File : VD067084.D
 Acq On : 05 Oct 2020 15:30
 Operator : VA/SY
 Sample : L4190-07
 Misc : 6.78G/5.00ml/MSVOA D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PAV-B14-092920-15-16

Quant Time: Oct 06 07:49:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 15:19:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	541038	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	891689	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	806216	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	362551	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	248647	47.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.42%	
35) Dibromofluoromethane	7.92	113	282477	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
50) Toluene-d8	10.34	98	1059468	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
62) 4-Bromofluorobenzene	12.64	95	339842	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
Target Compounds						
16) Acetone	4.16	43	7929	9.967	ug/l #	64
17) Carbon Disulfide	4.41	76	24913	1.592	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD100520\
Data File : VD067084.D
Acq On : 05 Oct 2020 15:30
Operator : VA/SY
Sample : L4190-07
Misc : 6.78G/5.00ml/MSVOA D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PAV-B14-092920-15-16

Quant Time: Oct 06 07:49:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 15:19:17 2020
Response via : Initial Calibration

