

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100720\  
 Data File : VD067149.D  
 Acq On : 07 Oct 2020 13:43  
 Operator : VA/SY  
 Sample : L4224-01  
 Misc : 10.51G/5.00ml/MSVOA D/SOIL  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 OR-02-100620

Quant Time: Oct 08 08:02:04 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_D\METHOD\82D100120S.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 01 15:19:17 2020  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	381280	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	603058	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	528325	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	242052	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	182780	49.77	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.54%	
35) Dibromofluoromethane	7.91	113	200801	50.86	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.72%	
50) Toluene-d8	10.34	98	690218	47.98	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.96%	
62) 4-Bromofluorobenzene	12.64	95	217899	44.62	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	89.24%	
Target Compounds						
16) Acetone	4.16	43	4097	7.308	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD100720\  
 Data File : VD067149.D  
 Acq On : 07 Oct 2020 13:43  
 Operator : VA/SY  
 Sample : L4224-01  
 Misc : 10.51G/5.00ml/MSVOA D/SOIL  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 OR-02-100620

Quant Time: Oct 08 08:02:04 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_D\METHOD\82D100120S.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 01 15:19:17 2020  
 Response via : Initial Calibration

